



# Analytical Report for Northwest Pipe

ASL Report #: R1877

Project ID: 682722.GW.05

**Attn: Gretchen Gee**

cc:

Beckett, Jamie/RDD

Authorized and Released By:

Laboratory Project Manager  
Kathy McKinley  
(541) 758-0235 ext.23144  
May 12, 2017

All analyses performed by CH2M HILL are clearly indicated. Any subcontracted analyses are included as appended reports as received from the subcontracted laboratory. The results included in this report only relate to the samples listed on the following Sample Cross-Reference page. This report shall not be reproduced except in full, without the written approval of the laboratory.

Any unusual difficulties encountered during the analysis of your samples are discussed in the attached case narratives.



Accredited in accordance with NELAP:  
Oregon (100022)  
Louisiana (05031)



ASL Report #: R1877

### Sample Receipt Comments

We certify that the test results meet all NELAP requirements.

### Sample Cross-Reference

<b>ASL Sample ID</b>	<b>Client Sample ID</b>	<b>Date/Time Collected</b>	<b>Date Received</b>
R187701	TRIP BLANK-042717-02	04/27/17 08:00	04/28/17
R187702	MW-02-042717-0	04/27/17 10:00	04/28/17
R187703	MW-04-042717-0	04/27/17 11:30	04/28/17
R187704	T4S1MW-23-042717-0	04/27/17 13:15	04/28/17
R187705	T4S1MW-22-042717	04/27/17 14:40	04/28/17

## CASE NARRATIVE GC/MS VOLATILES ANALYSIS

**Lab Name:** CH2M ASL

**ASL SDG#:** R1877

**Project:** Northwest Pipe

**Project #:** 682722.GW.05

---

With the exceptions noted as flags, footnotes, or detailed in the section below; standard operating procedures were followed in the analysis of the samples and no problems were encountered or anomalies observed.

All laboratory quality control samples were within established control limits, with any exceptions noted below, or in the associated QC summary forms.

Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. For diluted samples, the reporting limits are adjusted for the dilution required.

Calculations are performed before rounding to minimize errors in calculated values.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the section below, or in the sample receipt documentation.

**Method(s):**  
SW8260B: SW5030

# CH2M ASL

Client Information				Lab Information			
<b>Client Sample ID: TRIP BLANK-042717-02</b>				<b>Lab Sample ID: R187701</b>			
Project Name: Northwest Pipe				Date Received: 04/28/17			
Sample Date: 04/27/17				Dilution Factor: 1			
Sample Time: 08:00				Report Revision No.: 0			
Type: Grab							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
cis-1,2-Dichloroethene	156-59-2	0.15	0.50	0.15	U	ug/L	SW8260B	04/28/17
Trichloroethene (TCE)	79-01-6	0.15	0.50	0.15	U	ug/L	SW8260B	04/28/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	110	70-130	
1,2-Dichloroethane-d4	107	70-130	
Toluene-d8	99	70-130	
4-Bromofluorobenzene	97	70-130	

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
<b>Client Sample ID: MW-02-042717-0</b>				<b>Lab Sample ID: R187702</b>			
Project Name: Northwest Pipe				Date Received: 04/28/17			
Sample Date: 04/27/17				Dilution Factor: 1			
Sample Time: 10:00				Report Revision No.: 0			
Type: Grab							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
cis-1,2-Dichloroethene	156-59-2	0.15	0.50	0.21	J	ug/L	SW8260B	04/28/17
Trichloroethene (TCE)	79-01-6	0.15	0.50	0.15	U	ug/L	SW8260B	04/28/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	107	70-130	
1,2-Dichloroethane-d4	107	70-130	
Toluene-d8	101	70-130	
4-Bromofluorobenzene	101	70-130	

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
Client Sample ID: T4S1MW-23-042717-0				Lab Sample ID: R187704			
Project Name: Northwest Pipe				Date Received: 04/28/17			
Sample Date: 04/27/17				Dilution Factor: 1			
Sample Time: 13:15				Report Revision No.: 0			
Type: Grab							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
cis-1,2-Dichloroethene	156-59-2	0.15	0.50	0.42	J	ug/L	SW8260B	04/28/17
Trichloroethene (TCE)	79-01-6	0.15	0.50	0.39	J	ug/L	SW8260B	04/28/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	108	70-130	
1,2-Dichloroethane-d4	107	70-130	
Toluene-d8	99	70-130	
4-Bromofluorobenzene	100	70-130	

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
Client Sample ID: T4S1MW-22-042717				Lab Sample ID: R187705			
Project Name: Northwest Pipe				Date Received: 04/28/17			
Sample Date: 04/27/17				Dilution Factor: 1			
Sample Time: 14:40				Report Revision No.: 0			
Type: Grab							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
cis-1,2-Dichloroethene	156-59-2	0.15	0.50	2.11		ug/L	SW8260B	04/28/17
Trichloroethene (TCE)	79-01-6	0.15	0.50	3.56		ug/L	SW8260B	04/28/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	113	70-130	
1,2-Dichloroethane-d4	111	70-130	
Toluene-d8	102	70-130	
4-Bromofluorobenzene	102	70-130	

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative  
 B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
<b>Client Sample ID: WB1-0428</b>				<b>Lab Sample ID: WB1-0428</b>			
Project Name: Northwest Pipe				Date Received: N/A			
Sample Date: N/A				Dilution Factor: 1			
Sample Time: N/A				Report Revision No.: 0			
Type: QC							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
cis-1,2-Dichloroethene	156-59-2	0.15	0.50	0.15	U	ug/L	SW8260B	04/28/17
Trichloroethene (TCE)	79-01-6	0.15	0.50	0.15	U	ug/L	SW8260B	04/28/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	106	70-130	
1,2-Dichloroethane-d4	102	70-130	
Toluene-d8	101	70-130	
4-Bromofluorobenzene	102	70-130	

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information		Lab Information	
Project Name: Northwest Pipe		LCS ID: BS1W0428	
Type: QC		Report Revision No.: 0	
Matrix: Water		Dilution Factor: 1	

Analyte	CAS#	Spike Amount	Sample Result	Units	%Recovery	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>							
cis-1,2-Dichloroethene	156-59-2	20.0	18.8	ug/L	94	SW8260B	04/28/17
Trichloroethene (TCE)	79-01-6	20.0	19.9	ug/L	100	SW8260B	04/28/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	102	70-130	
1,2-Dichloroethane-d4	96	70-130	
Toluene-d8	99	70-130	
4-Bromofluorobenzene	103	70-130	

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

## CASE NARRATIVE GC/MS VOLATILES ANALYSIS

**Lab Name:** CH2M ASL

**ASL SDG#:** R1877

**Project:** Northwest Pipe

**Project #:** 682722.GW.05

---

With the exceptions noted as flags, footnotes, or detailed in the section below; standard operating procedures were followed in the analysis of the samples and no problems were encountered or anomalies observed.

All laboratory quality control samples were within established control limits, with any exceptions noted below, or in the associated QC summary forms.

Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. For diluted samples, the reporting limits are adjusted for the dilution required.

Calculations are performed before rounding to minimize errors in calculated values.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the section below, or in the sample receipt documentation.

**Method(s):**  
SW8260B: SW5030

# CH2M ASL

Client Information				Lab Information			
<b>Client Sample ID: MW-04-042717-0</b>				<b>Lab Sample ID: R187703</b>			
Project Name: Northwest Pipe				Date Received: 04/28/17			
Sample Date: 04/27/17				Dilution Factor: 1			
Sample Time: 11:30				Report Revision No.: 0			
Type: Grab							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
Vinyl Chloride	75-01-4	0.15	0.50	16.6		ug/L	SW8260B	04/28/17
cis-1,2-Dichloroethene	156-59-2	0.15	0.50	111	E	ug/L	SW8260B	04/28/17
Trichloroethene (TCE)	79-01-6	0.15	0.50	29.7		ug/L	SW8260B	04/28/17
Tetrachloroethene (PCE)	127-18-4	0.15	0.50	14.4		ug/L	SW8260B	04/28/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	110	70-130	
1,2-Dichloroethane-d4	105	70-130	
Toluene-d8	101	70-130	
4-Bromofluorobenzene	102	70-130	

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
Client Sample ID: MW-04-042717-0DL				Lab Sample ID: R187703DL			
Project Name: Northwest Pipe				Date Received: 04/28/17			
Sample Date: 04/27/17				Dilution Factor: 10			
Sample Time: 11:30				Report Revision No.: 0			
Type: Grab							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
Vinyl Chloride	75-01-4	1.50	5.00	16.4		ug/L	SW8260B	05/03/17
cis-1,2-Dichloroethene	156-59-2	1.50	5.00	116		ug/L	SW8260B	05/03/17
Trichloroethene (TCE)	79-01-6	1.50	5.00	31.3		ug/L	SW8260B	05/03/17
Tetrachloroethene (PCE)	127-18-4	1.50	5.00	14.7		ug/L	SW8260B	05/03/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	112	70-130	
1,2-Dichloroethane-d4	109	70-130	
Toluene-d8	103	70-130	
4-Bromofluorobenzene	99	70-130	

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative  
 B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
<b>Client Sample ID: WB1-0428</b>				<b>Lab Sample ID: WB1-0428</b>			
Project Name: Northwest Pipe				Date Received: N/A			
Sample Date: N/A				Dilution Factor: 1			
Sample Time: N/A				Report Revision No.: 0			
Type: QC							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
Vinyl Chloride	75-01-4	0.15	0.50	0.15	U	ug/L	SW8260B	04/28/17
cis-1,2-Dichloroethene	156-59-2	0.15	0.50	0.15	U	ug/L	SW8260B	04/28/17
Trichloroethene (TCE)	79-01-6	0.15	0.50	0.15	U	ug/L	SW8260B	04/28/17
Tetrachloroethene (PCE)	127-18-4	0.15	0.50	0.15	U	ug/L	SW8260B	04/28/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	106	70-130	
1,2-Dichloroethane-d4	102	70-130	
Toluene-d8	101	70-130	
4-Bromofluorobenzene	102	70-130	

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative  
 B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
<b>Client Sample ID: WB1-0503</b>				<b>Lab Sample ID: WB1-0503</b>			
Project Name: Northwest Pipe				Date Received: N/A			
Sample Date: N/A				Dilution Factor: 1			
Sample Time: N/A				Report Revision No.: 0			
Type: QC							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
Vinyl Chloride	75-01-4	0.15	0.50	0.15	U	ug/L	SW8260B	05/03/17
cis-1,2-Dichloroethene	156-59-2	0.15	0.50	0.15	U	ug/L	SW8260B	05/03/17
Trichloroethene (TCE)	79-01-6	0.15	0.50	0.15	U	ug/L	SW8260B	05/03/17
Tetrachloroethene (PCE)	127-18-4	0.15	0.50	0.15	U	ug/L	SW8260B	05/03/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	110	70-130	
1,2-Dichloroethane-d4	107	70-130	
Toluene-d8	103	70-130	
4-Bromofluorobenzene	102	70-130	

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information		Lab Information	
Project Name: Northwest Pipe		LCS ID: BS1W0428	
Type: QC		Report Revision No.: 0	
Matrix: Water		Dilution Factor: 1	

Analyte	CAS#	Spike Amount	Sample Result	Units	%Recovery	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>							
Vinyl Chloride	75-01-4	20.0	21.2	ug/L	106	SW8260B	04/28/17
cis-1,2-Dichloroethene	156-59-2	20.0	18.8	ug/L	94	SW8260B	04/28/17
Trichloroethene (TCE)	79-01-6	20.0	19.9	ug/L	100	SW8260B	04/28/17
Tetrachloroethene (PCE)	127-18-4	20.0	20.2	ug/L	101	SW8260B	04/28/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	102	70-130	
1,2-Dichloroethane-d4	96	70-130	
Toluene-d8	99	70-130	
4-Bromofluorobenzene	103	70-130	

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative

# CH2M ASL

Client Information		Lab Information	
Project Name: Northwest Pipe		LCS ID: BS1W0503	
Type: QC		Report Revision No.: 0	
Matrix: Water		Dilution Factor: 1	

Analyte	CAS#	Spike Amount	Sample Result	Units	%Recovery	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>							
Vinyl Chloride	75-01-4	20.0	20.1	ug/L	101	SW8260B	05/03/17
cis-1,2-Dichloroethene	156-59-2	20.0	17.4	ug/L	87	SW8260B	05/03/17
Trichloroethene (TCE)	79-01-6	20.0	18.5	ug/L	93	SW8260B	05/03/17
Tetrachloroethene (PCE)	127-18-4	20.0	18.3	ug/L	92	SW8260B	05/03/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	106	70-130	
1,2-Dichloroethane-d4	102	70-130	
Toluene-d8	102	70-130	
4-Bromofluorobenzene	104	70-130	

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative

## CASE NARRATIVE GC/MS VOLATILES ANALYSIS

**Lab Name:** CH2M ASL

**ASL SDG#:** R1877

**Project:** Northwest Pipe

**Project #:** 682722.GW.05

---

With the exceptions noted as flags, footnotes, or detailed in the section below; standard operating procedures were followed in the analysis of the samples and no problems were encountered or anomalies observed.

All laboratory quality control samples were within established control limits, with any exceptions noted below, or in the associated QC summary forms.

Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. For diluted samples, the reporting limits are adjusted for the dilution required.

Calculations are performed before rounding to minimize errors in calculated values.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the section below, or in the sample receipt documentation.

**Method(s):**

SW8260C-SIM: SW5030

# CH2M ASL

Client Information				Lab Information			
<b>Client Sample ID: TRIP BLANK-042717-02</b>				<b>Lab Sample ID: R187701</b>			
Project Name: Northwest Pipe				Date Received: 04/28/17			
Sample Date: 04/27/17				Dilution Factor: 1			
Sample Time: 08:00				Report Revision No.: 0			
Type: Grab							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
Vinyl Chloride	75-01-4	8.00	20.0	8.00	U	ng/L	SW8260C-SI	05/09/17
Tetrachloroethene (PCE)	127-18-4	5.00	20.0	5.00	U	ng/L	SW8260C-SI	05/09/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	106	70-130	
1,2-Dichloroethane-d4	105	70-130	
Toluene-d8	95	70-130	
4-Bromofluorobenzene	111	70-130	

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative  
 B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
<b>Client Sample ID: MW-02-042717-0</b>				<b>Lab Sample ID: R187702</b>			
Project Name: Northwest Pipe				Date Received: 04/28/17			
Sample Date: 04/27/17				Dilution Factor: 1			
Sample Time: 10:00				Report Revision No.: 0			
Type: Grab							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
Vinyl Chloride	75-01-4	8.00	20.0	11.5	J	ng/L	SW8260C-SI	05/09/17
Tetrachloroethene (PCE)	127-18-4	5.00	20.0	224		ng/L	SW8260C-SI	05/09/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	107	70-130	
1,2-Dichloroethane-d4	107	70-130	
Toluene-d8	94	70-130	
4-Bromofluorobenzene	109	70-130	

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
Client Sample ID: T4S1MW-23-042717-0				Lab Sample ID: R187704			
Project Name: Northwest Pipe				Date Received: 04/28/17			
Sample Date: 04/27/17				Dilution Factor: 1			
Sample Time: 13:15				Report Revision No.: 0			
Type: Grab							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
Vinyl Chloride	75-01-4	8.00	20.0	8.00	U	ng/L	SW8260C-SI	05/09/17
Tetrachloroethene (PCE)	127-18-4	5.00	20.0	1070		ng/L	SW8260C-SI	05/09/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	110	70-130	
1,2-Dichloroethane-d4	110	70-130	
Toluene-d8	95	70-130	
4-Bromofluorobenzene	111	70-130	

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative  
 B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
Client Sample ID: T4S1MW-22-042717				Lab Sample ID: R187705			
Project Name: Northwest Pipe				Date Received: 04/28/17			
Sample Date: 04/27/17				Dilution Factor: 1			
Sample Time: 14:40				Report Revision No.: 0			
Type: Grab							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
Vinyl Chloride	75-01-4	8.00	20.0	22.0		ng/L	SW8260C-SI	05/09/17
Tetrachloroethene (PCE)	127-18-4	5.00	20.0	1430		ng/L	SW8260C-SI	05/09/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	107	70-130	
1,2-Dichloroethane-d4	107	70-130	
Toluene-d8	93	70-130	
4-Bromofluorobenzene	110	70-130	

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative  
 B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
<b>Client Sample ID: WB1-0509</b>				<b>Lab Sample ID: WB1-0509</b>			
Project Name: Northwest Pipe				Date Received: N/A			
Sample Date: N/A				Dilution Factor: 1			
Sample Time: N/A				Report Revision No.: 0			
Type: QC							
Matrix: Water							

Analyte	CAS#	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>								
Vinyl Chloride	75-01-4	8.00	20.0	8.00	U	ng/L	SW8260C-SI	05/09/17
Tetrachloroethene (PCE)	127-18-4	5.00	20.0	5.00	U	ng/L	SW8260C-SI	05/09/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	103	70-130	
1,2-Dichloroethane-d4	102	70-130	
Toluene-d8	95	70-130	
4-Bromofluorobenzene	109	70-130	

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative  
 B=Analyte detected in blank

# CH2M ASL

Client Information		Lab Information	
Project Name: Northwest Pipe		LCS ID: BS1W0509	
Type: QC		Report Revision No.: 0	
Matrix: Water		Dilution Factor: 1	

Analyte	CAS#	Spike Amount	Sample Result	Units	%Recovery	Analysis Method	Date Analyzed
<b>GC/MS Volatiles</b>							
Vinyl Chloride	75-01-4	500	503	ng/L	101	SW8260C-SIM	05/09/17
Tetrachloroethene (PCE)	127-18-4	500	485	ng/L	97	SW8260C-SIM	05/09/17

Surrogate	% Recovery	Control Limits	Qualifier
Dibromofluoromethane	101	70-130	
1,2-Dichloroethane-d4	101	70-130	
Toluene-d8	98	70-130	
4-Bromofluorobenzene	111	70-130	

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative

## CASE NARRATIVE HEADSPACE ANALYSIS

**Lab Name:** CH2M ASL

**ASL SDG#:** R1877

**Project:** Northwest Pipe

**Project #:** 682722.GW.05

---

With the exceptions noted as flags, footnotes, or detailed in the section below; standard operating procedures were followed in the analysis of the samples and no problems were encountered or anomalies observed.

All laboratory quality control samples were within established control limits, with any exceptions noted below, or in the associated QC summary forms.

Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. For diluted samples, the reporting limits are adjusted for the dilution required.

Calculations are performed before rounding to minimize errors in calculated values.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the section below, or in the sample receipt documentation.

**Method(s):**  
RSK-175

# CH2M ASL

Client Information					Lab Information				
Client Sample ID: MW-02-042717-0					Lab Sample ID: R187702				
Project Name: Northwest Pipe					Date Received: 04/28/17				
Sample Date: 04/27/17					Report Revision No: 0				
Sample Time: 10:00									
Type: Grab									
Matrix: Water									

Analyte	CAS#	Dilution Factor	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC Volatiles</b>									
Methane	74-82-8	1	5.31	42.6	3420		ug/L	RSK-175	05/03/17
Carbon dioxide	124-38-9	1	37.2	225	15600		ug/L	RSK-175	05/03/17

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information					Lab Information				
Client Sample ID: MW-04-042717-0					Lab Sample ID: R187703				
Project Name: Northwest Pipe					Date Received: 04/28/17				
Sample Date: 04/27/17					Report Revision No: 0				
Sample Time: 11:30									
Type: Grab									
Matrix: Water									

Analyte	CAS#	Dilution Factor	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC Volatiles</b>									
Methane	74-82-8	1	4.67	37.5	1210		ug/L	RSK-175	05/03/17
Carbon dioxide	124-38-9	1	35.8	217	82000		ug/L	RSK-175	05/03/17

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information	Lab Information
Client Sample ID: T4S1MW-23-042717-0	Lab Sample ID: R187704
Project Name: Northwest Pipe	Date Received: 04/28/17
Sample Date: 04/27/17	Report Revision No: 0
Sample Time: 13:15	
Type: Grab	
Matrix: Water	

Analyte	CAS#	Dilution Factor	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC Volatiles</b>									
Methane	74-82-8	1	5.86	47.0	7.12	J	ug/L	RSK-175	05/03/17
Carbon dioxide	124-38-9	1	38.4	232	20400		ug/L	RSK-175	05/03/17

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information					Lab Information				
Client Sample ID: T4S1MW-22-042717					Lab Sample ID: R187705				
Project Name: Northwest Pipe					Date Received: 04/28/17				
Sample Date: 04/27/17					Report Revision No: 0				
Sample Time: 14:40									
Type: Grab									
Matrix: Water									

Analyte	CAS#	Dilution Factor	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC Volatiles</b>									
Methane	74-82-8	1	6.05	48.6	6.05	U	ug/L	RSK-175	05/03/17
Carbon dioxide	124-38-9	1	38.8	235	49100		ug/L	RSK-175	05/03/17

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information					Lab Information				
Project Name: Northwest Pipe					Method Blank ID: XB1-0503				
Sample Date: N/A					Date Received: N/A				
Sample Time: N/A					Report Revision No: 0				
Type: QC									
Matrix: Water									

Analyte	CAS#	Dilution Factor	DL	RL	Sample Result	Qualifier	Units	Analysis Method	Date Analyzed
<b>GC Volatiles</b>									
Methane	74-82-8	1	14.3	114	14.3	U	ug/L	RSK-175	05/03/17
Carbon dioxide	124-38-9	1	56.6	342	56.6	U	ug/L	RSK-175	05/03/17

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
Project Name: Northwest Pipe Type: QC Matrix: Water				LCS ID: BS1X0503 Report Revision No.: 0 Dilution Factor: 1			

Analyte	CAS#	Spike Amount	Sample Result	Units	%Recovery	Analysis Method	Date Analyzed
<b>GC Volatiles</b>							
Methane	74-82-8	593	594	ug/L	100	RSK-175	05/03/17
Carbon dioxide	124-38-9	3180	3140	ug/L	99	RSK-175	05/03/17

U=Not detected and reported as less than detection limit  
J=Estimated value below reporting limit  
E=Estimated value above calibration range  
\*=See case narrative

## CASE NARRATIVE METALS ANALYSIS

**Lab Name:** CH2M ASL

**ASL SDG#:** R1877

**Project:** Northwest Pipe

**Project #:** 682722.GW.05

---

With the exceptions noted as flags, footnotes, or detailed in the section below; standard operating procedures were followed in the analysis of the samples and no problems were encountered or anomalies observed.

All laboratory quality control samples were within established control limits, with any exceptions noted below, or in the associated QC summary forms.

Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. For diluted samples, the reporting limits are adjusted for the dilution required.

Calculations are performed before rounding to minimize errors in calculated values.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the section below, or in the sample receipt documentation.

**Method(s):**

E200.7: FLDFLT

# CH2M ASL

Client Information				Lab Information			
Project Name: Northwest Pipe				Lab Batch ID: R1877			
Date Received: 04/28/17				Report Revision No.: 0			
Type: See C.O.C.							
Matrix: Water							

Client Sample ID	Lab Sample ID	Dilution Factor	DL	RL	Result	Qual	Units	Date Analyzed
<b>Iron: E200.7</b>								
<i>Dissolved Metals</i>								
MW-02-042717-0	R187702F	1	10.0	100	1340		ug/L	05/05/17
MW-04-042717-0	R187703F	1	10.0	100	9830		ug/L	05/05/17
T4S1MW-23-042717-0	R187704F	1	10.0	100	54.5	J	ug/L	05/05/17
T4S1MW-22-042717	R187705F	1	10.0	100	10.0	U	ug/L	05/05/17
<i>Total Metals</i>								
WB10-0505	WB10-0505	1	10.0	100	10.0	U	ug/L	05/05/17

U=Not detected and reported as less than detection limit

J=Estimated value below reporting limit

E=Estimated value above calibration range

\*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information	Lab Information
Project Name: Northwest Pipe Type: QC Matrix: Water	Blank Spike ID: BS10W0505 Report Revision No: 0 Dilution Factor: 1

Analyte	Spike Amount	Result	Units	%Recovery	Analysis Method	Prep Method	Date Analyzed
<b>Metals</b>							
Iron	50000	52700	ug/L	105	E200.7	E200.2	05/05/17

U=Not detected and report as less than detection limit  
J=Estimated value below reporting limit  
E=Estimated value above calibration range  
\*=See case narrative

## CASE NARRATIVE GENERAL CHEMISTRY ANALYSIS

**Lab Name:** CH2M ASL

**ASL SDG#:** R1877

**Project:** Northwest Pipe

**Project #:** 682722.GW.05

---

With the exceptions noted as flags, footnotes, or detailed in the section below; standard operating procedures were followed in the analysis of the samples and no problems were encountered or anomalies observed.

All laboratory quality control samples were within established control limits, with any exceptions noted below, or in the associated QC summary forms.

Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. For diluted samples, the reporting limits are adjusted for the dilution required.

Calculations are performed before rounding to minimize errors in calculated values.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the section below, or in the sample receipt documentation.

**Method(s):**  
E300.0A

# CH2M ASL

Client Information			Lab Information		
<b>Project Name: Northwest Pipe</b>			<b>Lab Batch ID: R1877</b>		
Date Received: 04/28/2017			Analysis Method: E300.0A		
Type: See C.O.C.			Units: mg/L		
Matrix: Water			Report Revision No.: 0		

Client Sample ID	Lab Sample ID	Dilution Factor	DL	Chloride RL	Result	Qualifier	Date Analyzed
<b>General Chemistry</b>							
MW-02-042717-0	R187702	1	0.020	0.20	2.47		05/04/2017
MW-04-042717-0	R187703	1	0.020	0.20	4.52		05/04/2017
T4S1MW-23-042717-0	R187704	1	0.020	0.20	3.36		05/04/2017
T4S1MW-22-042717	R187705	1	0.020	0.20	3.38		05/04/2017
WB1-0504	WB1-0504	1	0.020	0.20	0.020	U	05/04/2017

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information			Lab Information		
<b>Project Name: Northwest Pipe</b>			<b>Lab Batch ID: R1877</b>		
Date Received: 04/28/2017			Analysis Method: E300.0A		
Type: See C.O.C.			Units: mg/L		
Matrix: Water			Report Revision No.: 0		

Client Sample ID	Lab Sample ID	Dilution Factor	DL	RL	Sulfate Result	Qualifier	Date Analyzed
<b>General Chemistry</b>							
MW-02-042717-0	R187702	1	0.020	0.20	5.37		05/04/2017
MW-04-042717-0	R187703	1	0.020	0.20	2.35		05/04/2017
T4S1MW-23-042717-0	R187704	1	0.020	0.20	7.76		05/04/2017
T4S1MW-22-042717	R187705	1	0.020	0.20	10.0		05/04/2017
WB1-0504	WB1-0504	1	0.020	0.20	0.020	U	05/04/2017

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
Project Name: Northwest Pipe Type: QC Matrix: Water				Lab Batch ID: R1877 Report Revision No.: 0			

LCS ID	Analyte	Spike Amount	Sample Result	Units	% Recovery	Analysis Method	Date Analyzed
<b>General Chemistry</b>							
BS1W0504	Chloride	5.00	4.99	mg/L	100	E300.0A	05/04/2017
BS1W0504	Sulfate	5.00	4.85	mg/L	97	E300.0A	05/04/2017

U=Not detected and reported as less than detection limit  
J=Estimated value below reporting limit  
E=Estimated value above calibration range  
\*=See case narrative

## CASE NARRATIVE GENERAL CHEMISTRY ANALYSIS

**Lab Name:** CH2M ASL

**ASL SDG#:** R1877

**Project:** Northwest Pipe

**Project #:** 682722.GW.05

---

With the exceptions noted as flags, footnotes, or detailed in the section below; standard operating procedures were followed in the analysis of the samples and no problems were encountered or anomalies observed.

All laboratory quality control samples were within established control limits, with any exceptions noted below, or in the associated QC summary forms.

Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. For diluted samples, the reporting limits are adjusted for the dilution required.

Calculations are performed before rounding to minimize errors in calculated values.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the section below, or in the sample receipt documentation.

**Method(s):**

E353.2

# CH2M ASL

Client Information			Lab Information		
Project Name: Northwest Pipe			Lab Batch ID: R1877		
Date Received: 04/28/17			Analysis Method: E353.2		
Type: See C.O.C.			Units: mg/L		
Matrix: Water			Report Revision No.: 0		

Client Sample ID	Lab Sample ID	Dilution Factor	DL	Nitrate-N RL	Result	Qualifier	Date Analyzed
<b>General Chemistry</b>							
MW-02-042717-0	R187702	1	0.0028	0.010	0.31		04/28/17 14:17
MW-04-042717-0	R187703	1	0.0028	0.010	0.011		04/28/17 14:18
T4S1MW-23-042717-0	R187704	1	0.0028	0.010	0.42		04/28/17 14:20
T4S1MW-22-042717	R187705	1	0.0028	0.010	0.92		04/28/17 14:21
WB1-042817	WB1-042817	1	0.0028	0.010	0.0028	U	04/28/17

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative

B=Analyte detected in blank

## CASE NARRATIVE GENERAL CHEMISTRY ANALYSIS

**Lab Name:** CH2M ASL

**ASL SDG#:** R1877

**Project:** Northwest Pipe

**Project #:** 682722.GW.05

---

With the exceptions noted as flags, footnotes, or detailed in the section below; standard operating procedures were followed in the analysis of the samples and no problems were encountered or anomalies observed.

All laboratory quality control samples were within established control limits, with any exceptions noted below, or in the associated QC summary forms.

Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. For diluted samples, the reporting limits are adjusted for the dilution required.

Calculations are performed before rounding to minimize errors in calculated values.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the section below, or in the sample receipt documentation.

**Method(s):**  
SM5310B

# CH2M ASL

Client Information			Lab Information			
<b>Project Name: Northwest Pipe</b>			<b>Lab Batch ID: R1877</b>			
Date Received: 04/28/2017			Analysis Method: SM5310B			
Type: See C.O.C.			Units: mg/L			
Matrix: Water			Report Revision No.: 0			

Client Sample ID	Lab Sample ID	Dilution Factor	DL	Total Organic Carbon RL	Result	Qualifier	Date Analyzed
<b>General Chemistry</b>							
MW-02-042717-0	R187702	1	0.20	0.50	1.60		05/08/2017
MW-04-042717-0	R187703	1	0.20	0.50	1.40		05/08/2017
T4S1MW-23-042717-0	R187704	1	0.20	0.50	0.78		05/08/2017
T4S1MW-22-042717	R187705	1	0.20	0.50	1.18		05/08/2017
WB1-0508	WB1-0508	1	0.20	0.50	0.20	U	05/08/2017

U=Not detected and reported as less than detection limit  
 J=Estimated value below reporting limit  
 E=Estimated value above calibration range  
 \*=See case narrative

B=Analyte detected in blank

# CH2M ASL

Client Information				Lab Information			
Project Name: Northwest Pipe Type: QC Matrix: Water				Lab Batch ID: R1877 Report Revision No.: 0			

LCS ID	Analyte	Spike Amount	Sample Result	Units	% Recovery	Analysis Method	Date Analyzed
<b>General Chemistry</b>							
BS1W0508	Total Organic Carbon	5.00	4.87	mg/L	97	SM5310B	05/08/2017

U=Not detected and reported as less than detection limit  
J=Estimated value below reporting limit  
E=Estimated value above calibration range  
\*=See case narrative

# Chain of Custody Record

<b>Client Contact</b>	<b>Analysis Turnaround Time</b>	<b>Preservation Used</b>	<b>For Lab Use Only:</b>
Project Name: <b>NWP</b>	TAT is Calander days	4 1 1 3 4 4 1 4	SDG: <b>R1877</b>
Project # or PO #: <b>602722</b>	TAT if different from below _____	<b>Analysis Requested</b>	Custody Seals intact? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
Company Name: <b>CH2M HILL</b>	<input checked="" type="checkbox"/> 21 days (STD)	E200.7 F	Hand delivered? <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No
Address: <b>2020 SW 4th Ave Ste 300</b>	<input type="checkbox"/> 14 days * <input type="checkbox"/> 3 day *	E300.0, E353.2	Cooler Temp: <b>3.5</b> °C
City/State/Zip: <b>Portland, OR 97201</b>	<input type="checkbox"/> 7 days * <input type="checkbox"/> 2 days *	RSK-175	Therm ID No. <b>173</b> Therm Exp. <b>7/17/17</b>
Project Manager: <b>Gretchen Gee</b>	<input type="checkbox"/> 5 days * <input type="checkbox"/> 1 day *	SM 5310	Packing Material: Circle Below
Phone #: <b>503 736 4349</b>	(Surcharges will apply)	SW82600	<input checked="" type="radio"/> Blue Ice <input type="radio"/> Ice Box <input type="radio"/> Bubble Wrap
Report to email: <b>Gretchen.Gee@ch2m.com</b>		SW82600-1M	Radiological Screen? <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No

Sample Identification (Limit of 20 characters)	Sample Date	Sample Time	Sample Type (CuComp, G=Grab)	Matrix (Water, Soil, Air)	Total # of Cont.	E200.7 F	E300.0, E353.2	RSK-175	SM 5310	SW82600	SW82600-1M	E300.0, E353.2, E200.7*	E200.7 F	Sample Specific Notes:	Lab ID:
TRIPBLANK-042717-02	4/27/17	0800	G	W	4										1
MW-02-042717-0	4/27/17	1000	G	W	12			3	1	3	3	1			2
MW-04-042717-0	4/27/17	1130	G	W	12			3	1	3	3	1			3
T4S1MW-23-042717-0	4/27/17	1315	G	W	12			3	1	3	3	1			4
T4S1MW-22-042717-0	4/27/17	1440	G	W	12	1	1	3	1	3	3				5

Preservation Used: 1= Ice, 2= HCl; 3= H2SO4; 4=HNO3; 5=NaOH; 6= Other

<b>Possible Hazard Identification:</b>	<b>Sample Disposal</b> (A fee may be added if samples are retained longer than 30 day per client request, samples are returned to client, or classified as hazardous.)
Are samples hazardous? <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	
If yes, select hazard(s): <input type="checkbox"/> Listed <input type="checkbox"/> Ignitable <input type="checkbox"/> Corrosive <input type="checkbox"/> Reactive <input type="checkbox"/> Toxic	
If YES or NO is not checked above, samples will be assumed hazardous and hazardous disposal fees will be applied.	<input type="checkbox"/> Return to Client <input checked="" type="checkbox"/> Disposal by Lab <input type="checkbox"/> Archive for _____ months

Sampled by: <b>Jennifer Ulrich</b>	Date/Time: <b>4/27/17</b>	Relinquished by: <b>Jennifer Ulrich</b>	Date/Time: <b>4/27/17/1615</b>
Received by:	Date/Time:	Relinquished by:	Date/Time:
Received in Laboratory by: <b>Tina Williams</b>	Date/Time: <b>4/28/17 1015</b>	Shipped Via: <input type="checkbox"/> UPS <input checked="" type="checkbox"/> Fed-Ex <input type="checkbox"/> USPS <input type="checkbox"/> Other	Tracking #:

Special Instructions/QC Requirements  
 • E200.7 F = field filtered; filter E200.7 from un-preserved group E300.0, E353.2, E200.7

## Sample Receipt Record

SDG ID: R1877

Date Received: 4/28/2017

Client/Project: Northwest Pipe

Received by: TW

Were custody seals intact and on the outside of the cooler?	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A
Shipping Record:	<input type="checkbox"/> Hand Delivered	<input checked="" type="checkbox"/> On File	<input type="checkbox"/> COC
Radiological Screening for DoD	<input type="checkbox"/> Yes	<input type="checkbox"/> No	<input checked="" type="checkbox"/> N/A
Packing Material:	<input type="checkbox"/> Hand Delivered	<input checked="" type="checkbox"/> Ice	<input type="checkbox"/> Blue Ice <input type="checkbox"/> Box
Temp OK? (<6C) Therm ID: TH173 Exp. 7/17/17	3.5 °C	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No <input type="checkbox"/> N/A
Was a Chain of Custody (CoC) Provided?	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A
Was the CoC correctly filled out (If No, document below)	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A
Did sample labels agree with COC? (If No, document below)	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A
Did the CoC list a correct bottle count and the preservative types (No=Correct on CoC)	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A
Were the sample containers in good condition (not broken or leaking)?	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A
Was enough sample volume provided for analysis? (If No, document below)	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A
Containers supplied by ASL?	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A
Any sample with < 1/2 holding time remaining? If so contact LPM and document below.	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A
Samples have multi-phase? If yes, document on SRER	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No	<input type="checkbox"/> N/A
All water VOCs free of air bubbles? No, document on SRER	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A
pH of all samples met criteria on receipt? If "No", preserve and document below.	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	<input type="checkbox"/> N/A
Dissolved/Soluble metals filtered in the field?	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No	<input type="checkbox"/> N/A
Dissolved/Soluble metals have sediment in bottom of container? If so document below.	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No	<input type="checkbox"/> N/A

### Preservation Adjustment

Sample ID	Reagent	Reagent Lot Number	Volume Added	Initials/Date-Time	24 hour pH check Initials/Time

Did pH of all metals samples preserved upon receipt meet criteria 24 hours after preservation?  Yes  No

### Sample Exception Report (The following exceptions were noted)

- 1) MW-02-42717-0 (R187702) Nitrate sample has less than half holding time remaining.
- 2) Samples MW-02-042717-0, MW-04-042717-0, and T4S1MW-23-042717-0 (R1877-02, -03, and -04) were filtered into 8oz poly bottles preserved with HNO3 (Lot #937) using 0.45um Sterivex filters (Lot #1616/00279).
- 3) Compound list not included on COC. Logged in according to quote and past SDGs.

Client was notified on: 4/28/17 Client contact: Jamie Beckett/RDD

Resolution to Exception: