

10/19/2021

Mr. Lucas Henry

GeoSyntec Consultants

920 SW Sixth Avenue, Suite 600

Portland OR 97204

Project Name: Cascade SVE

Project #: 0564519

Workorder #: 2110145

Dear Mr. Lucas Henry

The following report includes the data for the above referenced project for sample(s) received on 10/6/2021 at Eurofins Air Toxics LLC.

The data and associated QC analyzed by Modified TO-15 are compliant with the project requirements or laboratory criteria with the exception of the deviations noted in the attached case narrative.

Thank you for choosing Eurofins Air Toxics LLC. for your air analysis needs. Eurofins Air Toxics Inc. is committed to providing accurate data of the highest quality. Please feel free to contact the Project Manager: Monica Tran at 916-985-1000 if you have any questions regarding the data in this report.

Regards,



Monica Tran

Project Manager

WORK ORDER #: 2110145

Work Order Summary

CLIENT:	Mr. Lucas Henry GeoSyntec Consultants 920 SW Sixth Avenue, Suite 600 Portland, OR 97204	BILL TO:	Accounts Payable Cascade Corporation 2201 NE 201st Avenue Fairview, OR 97024
PHONE:	503-222-9518	P.O. #	
FAX:	503-242-1416	PROJECT #	0564519 Cascade SVE
DATE RECEIVED:	10/06/2021	CONTACT:	Monica Tran
DATE COMPLETED:	10/19/2021		

<u>FRACTION #</u>	<u>NAME</u>	<u>TEST</u>	<u>RECEIPT VAC./PRES.</u>	<u>FINAL PRESSURE</u>
01A	SVE-EFF-100521	Modified TO-15	5.7 "Hg	10 psi
02A	Lab Blank	Modified TO-15	NA	NA
03A	CCV	Modified TO-15	NA	NA
04A	LCS	Modified TO-15	NA	NA
04AA	LCSD	Modified TO-15	NA	NA

CERTIFIED BY: 

 Technical Director

DATE: 10/19/21

Certification numbers: AZ Licensure AZ0775, FL NELAP – E87680, LA NELAP – 02089, NH NELAP - 209220, NJ NELAP - CA016, NY NELAP - 11291, TX NELAP - T104704434-20-16, UT NELAP – CA009332020-12, VA NELAP - 10615, WA NELAP - C935

Name of Accreditation Body: NELAP/ORELAP (Oregon Environmental Laboratory Accreditation Program)

Accreditation number: CA300005-014, Effective date: 10/18/2020, Expiration date: 10/17/2021.

Eurofins Air Toxics, LLC certifies that the test results contained in this report meet all requirements of the NELAC standards

This report shall not be reproduced, except in full, without the written approval of Eurofins Air Toxics, LLC.

180 BLUE RAVINE ROAD, SUITE B FOLSOM, CA - 95630
 (916) 985-1000 . (800) 985-5955 . FAX (916) 351-8279

LABORATORY NARRATIVE
Modified TO-15
GeoSyntec Consultants
Workorder# 2110145

One 1 Liter Summa Canister (100% Certified) sample was received on October 06, 2021. The laboratory performed analysis via modified EPA Method TO-15 using GC/MS in the full scan mode.

Method modifications taken to run these samples are summarized in the table below. Specific project requirements may over-ride the EATL modifications.

<i>Requirement</i>	<i>TO-15</i>	<i>ATL Modifications</i>
Initial Calibration	</=30% RSD with 2 compounds allowed out to < 40% RSD	</=30% RSD with 4 compounds allowed out to < 40% RSD
Blank and standards	Zero Air	UHP Nitrogen provides a higher purity gas matrix than zero air

Receiving Notes

There were no receiving discrepancies.

Analytical Notes

There were no analytical discrepancies.

Definition of Data Qualifying Flags

Eight qualifiers may have been used on the data analysis sheets and indicates as follows:

B - Compound present in laboratory blank greater than reporting limit (background subtraction not performed).

J - Estimated value.

E - Exceeds instrument calibration range.

S - Saturated peak.

Q - Exceeds quality control limits.

U - Compound analyzed for but not detected above the reporting limit, LOD, or MDL value. See data page for project specific U-flag definition.

UJ- Non-detected compound associated with low bias in the CCV

N - The identification is based on presumptive evidence.

File extensions may have been used on the data analysis sheets and indicates as follows:

a-File was requantified

b-File was quantified by a second column and detector

r1-File was requantified for the purpose of reissue

Summary of Detected Compounds
MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

Client Sample ID: SVE-EFF-100521

Lab ID#: 2110145-01A

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
cis-1,2-Dichloroethene	0.69	14	2.7	56
Trichloroethene	0.69	140	3.7	760
Tetrachloroethene	0.69	8.7	4.7	59

Client Sample ID: SVE-EFF-100521

Lab ID#: 2110145-01A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	20101520	Date of Collection: 10/5/21 6:00:00 PM
Dil. Factor:	6.93	Date of Analysis: 10/15/21 08:59 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	0.69	Not Detected	1.8	Not Detected
1,1-Dichloroethene	0.69	Not Detected	2.7	Not Detected
cis-1,2-Dichloroethene	0.69	14	2.7	56
Trichloroethene	0.69	140	3.7	760
Tetrachloroethene	0.69	8.7	4.7	59

Container Type: 1 Liter Summa Canister (100% Certified)

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	115	70-130
Toluene-d8	129	70-130
4-Bromofluorobenzene	84	70-130

Client Sample ID: Lab Blank

Lab ID#: 2110145-02A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	20101506c	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 10/15/21 09:58 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	0.10	Not Detected	0.26	Not Detected
1,1-Dichloroethene	0.10	Not Detected	0.40	Not Detected
cis-1,2-Dichloroethene	0.10	Not Detected	0.40	Not Detected
Trichloroethene	0.10	Not Detected	0.54	Not Detected
Tetrachloroethene	0.10	Not Detected	0.68	Not Detected

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	117	70-130
Toluene-d8	122	70-130
4-Bromofluorobenzene	90	70-130

Client Sample ID: CCV

Lab ID#: 2110145-03A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	20101502	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 10/15/21 07:15 AM

Compound	%Recovery
Vinyl Chloride	107
1,1-Dichloroethene	104
cis-1,2-Dichloroethene	107
Trichloroethene	110
Tetrachloroethene	106

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	104	70-130
Toluene-d8	126	70-130
4-Bromofluorobenzene	111	70-130

Client Sample ID: LCS

Lab ID#: 2110145-04A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	20101503	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 10/15/21 08:01 AM

Compound	%Recovery	Method Limits
Vinyl Chloride	109	70-130
1,1-Dichloroethene	107	70-130
cis-1,2-Dichloroethene	108	70-130
Trichloroethene	108	70-130
Tetrachloroethene	106	70-130

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	101	70-130
Toluene-d8	120	70-130
4-Bromofluorobenzene	111	70-130

Client Sample ID: LCSD

Lab ID#: 2110145-04AA

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	20101504	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 10/15/21 08:40 AM

Compound	%Recovery	Method Limits
Vinyl Chloride	110	70-130
1,1-Dichloroethene	106	70-130
cis-1,2-Dichloroethene	108	70-130
Trichloroethene	105	70-130
Tetrachloroethene	104	70-130

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	102	70-130
Toluene-d8	115	70-130
4-Bromofluorobenzene	110	70-130